

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: CviQI
Catalog Number: R0639S
Concentration: 10,000 U/mI

Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg

of Lambda DNA in 1 hour at 25°C in a total reaction volume of 50 μl.

Packaging Lot Number: 1017125.
Expiration Date: 08/2024
Storage Temperature: -20°C

Storage Conditions: 250 mM NaCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50%

Glycerol, 0.15% Triton X-100, 200 μg/ml BSA

Specification Version: PS-R0639S/L v1.0

CviQI Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
R0639SVIAL	CviQI	10161185	Pass	
B6003SVIAL	NEBuffer™ r3.1	10146827	Pass	

Assay Name/Specification	Lot # 10171252
Ligation and Recutting (Terminal Integrity) After a 10-fold over-digestion of Lambda DNA with CviQI, ~50% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with CviQI.	Pass
Non-Specific DNase Activity (16 Hour) A 50 µl reaction in NEBuffer 3.1 containing 1 µg of Lambda DNA and a minimum of 10 units of CviQl incubated for 16 hours at 25°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
Exonuclease Activity (Radioactivity Release) A 50 μl reaction in NEBuffer 3.1 containing 1 μg of a mixture of single and double-stranded [³H] E. coli DNA and a minimum of 50 units of CviQI incubated for 4 hours at 25°C releases <0.1% of the total radioactivity.	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



R0639S / Lot: 10171252


Page 1 of 2



Stephanie Cornelio **Production Scientist** 10 Aug 2022

Michael Tonello

Packaging Quality Control Inspector 08 Nov 2022



Page 2 of 2