

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Certificate of Analysis

Product Name: BsmBl-v2
Catalog Number: R0739L
Concentration: 10,000 U/ml

Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg

of Lambda DNA in 1 hour at 55°C in a total reaction volume of 50 μl.

Packaging Lot Number: 10076651
Expiration Date: 06/2022
Storage Temperature: -20°C

Storage Conditions: 300 mM NaCl , 10 mM Tris-HCl , 1 mM DTT , 0.1 mM EDTA , 50 %

Glycerol , 500  $\mu$ g/ml BSA, (pH 7.4 @ 25°C)

Specification Version: PS-R0739S/L v1.0

BsmBI-v2 Component List				
<b>NEB Part Number</b>	Component Description	Lot Number	Individual QC Result	
R0739LVIAL	BsmBI-v2	10076652	Pass	
B7203SVIAL	NEBuffer™ 3.1	10072160	Pass	
B7024SVIAL	Gel Loading Dye, Purple (6X)	10071081	Pass	

Assay Name/Specification	Lot # 10076651
Functional Testing (15 minute Digest)  A 50 μl reaction in NEBuffer 3.1 containing 1 μg of Lambda DNA and 1 μl of BsmBl-v2	Pass
incubated for 15 minutes at 55°C results in complete digestion as determined by agarose gel electrophoresis.	
Protein Purity Assay (SDS-PAGE) BsmBI-v2 is ≥ 95% pure as determined by SDS-PAGE analysis using Coomassie Blue detection.	Pass
<b>Ligation and Recutting (Terminal Integrity)</b> After a 10-fold over-digestion of Lambda DNA with BsmBI-v2, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with BsmBI-v2.	Pass
Exonuclease Activity (Radioactivity Release) A 50 µl reaction in NEBuffer 3.1 containing 1 µg of a mixture of single and double-stranded [ ³H] E. coli DNA and a minimum of 50 units of BsmBl-v2 incubated for 4 hours at 55°C releases <0.1% of the total radioactivity.	Pass



R0739L / Lot: 10076651

Page 1 of 2

This product has been tested and shown to be in compliance with all specifications.

conversion to the nicked form as determined by agarose gel electrophoresis.

JianYing Luo Production Scientist 09 Jun 2020 Michael Tonello

Packaging Quality Control Inspector

09 Jun 2020

