

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name:BsmBI-v2Catalog Number:R0739SConcentration:10,000 U/ml

Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg

of Lambda DNA in 1 hour at 55°C in a total reaction volume of 50 μl.

Packaging Lot Number: 10069104
Expiration Date: 02/2022
Storage Temperature: -20°C

Storage Conditions: 300 mM NaCl, 10 mM Tris-HCl, 1 mM DTT, 0.1 mM EDTA, 50 %

Glycerol , 500 μg/ml BSA, (pH 7.4 @ 25°C)

Specification Version: PS-R0739S/L v1.0

BsmBI-v2 Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
R0739SVIAL	BsmBI-v2	10069103	Pass	
B7203SVIAL	NEBuffer™ 3.1	10053974	Pass	
B7024SVIAL	Gel Loading Dye, Purple (6X)	10059230	Pass	

Assay Name/Specification	Lot # 10069104
Protein Purity Assay (SDS-PAGE) BsmBl-v2 is ≥ 95% pure as determined by SDS-PAGE analysis using Coomassie Blue detection.	Pass
Non-Specific DNase Activity (16 Hour) A 50 µl reaction in NEBuffer 3.1 containing 1 µg of Lambda DNA and a minimum of 10 units of BsmBl-v2 incubated for 16 hours at 55°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
Ligation and Recutting (Terminal Integrity) After a 10-fold over-digestion of Lambda DNA with BsmBI-v2, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with BsmBI-v2.	Pass
Functional Testing (15 minute Digest) A 50 μl reaction in NEBuffer 3.1 containing 1 μg of Lambda DNA and 1 μl of BsmBl-v2 incubated for 15 minutes at 55°C results in complete digestion as determined by agarose gel electrophoresis.	Pass



R0739S / Lot: 10069104

Page 1 of 2

This product has been tested and shown to be in compliance with all specifications.

conversion to the nicked form as determined by agarose gel electrophoresis.

Jianying Luo Production Scientist

27 Feb 2020

Jay Minichiello

Packaging Quality Control Inspector

28 Feb 2020

