

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: Goat Anti-Rat IgG Magnetic Beads

Catalog Number: \$1433\$ Concentration: 20 mg/ml

Unit Definition:

Lot Number: 10017555
Expiration Date: 07/2021
Storage Temperature: 4°C

Storage Conditions: 0.02 % NaN3, 0.05 % Tween® 20, 0.1 % BSA, 1 X PBS, (pH 7.4 @ 25°C)

Specification Version: PS-S1433S v1.0

Goat Anti-Rat IgG Magnetic Beads Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
S1433SVIAL	Goat Anti-Rat IgG Magnetic Beads	10015067	Pass	

Assay Name/Specification	Lot # 10017555
Non-Specific DNase Activity (16 hour, Buffer) A 50 μl reaction in Goat Anti-Rat IgG Magnetic Bead Storage Buffer containing 1 μg of PhiX174-HaeIII DNA incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
RNase Activity (Buffer) A 10 µl reaction in Goat Anti-Rat IgG Magnetic Bead Storage Buffer containing 40 ng of a 300 base single-stranded RNA is incubated at 37°C. After incubation for 16 hours, >90% of the substrate RNA remains intact as determined by fluorescent detection.	Pass
Functional Binding Assay (Qualitative) Goat Anti-Rat IgG Magnetic Beads (1 mg) were equilibrated and incubated with 500 µl of Serum containing Rat IgG for 1 hour at 25°C, then washed, eluted and evaluated by Tris-Glycine gel to confirm low non-specific binding of extract proteins and high isolation of target.	Pass
Binding Capacity (Magnetic Beads) Goat Anti-Rat IgG Magnetic Beads (1 mg) were equilibrated and incubated with 200 µI of >0.5 mg/ml Purified Rat IgG for 1 hour at 25°C, then washed and the Rat IgG eluted. Binding capacity was determined to be >5 µg of Rat IgG per mg of beads.	Pass

This product has been tested and shown to be in compliance with all specifications.



S1433S / Lot: 10017555

Page 1 of 2



Brad Landgraf
Production Scientist

27 Jul 2018

Michael Tonello

Packaging Quality Control Inspector

09 Aug 2018