NEW ENGLAND BioLabs Inc.
be INSPIRED drive DISCOVERY stay GENUINE
240 County Road
Ipswich, MA 01938-2723
Tel 978-927-5054
Fax 978-921-1350
www.neb.com
info@neb.com

New England Biolabs
Certificate of Analysis

Product Name:	*Magnetic mRNA Isolation Kit*
Catalog Number:	*S1550S*
Packaging Lot Number:	*10079682*
Expiration Date:	*07/2022*
Storage Temperature:	*4°C*
Specification Version:	*PS-S1550S v2.0*

Magnetic mRNA Isolation Kit Component List			
NEB Part Number	**Component Description**	**Lot Number**	**Individual QC Result**

Assay Name/Specification	**Lot # 10079682**
*** Individual Product Component Note** Standard Quality Control Tests are performed for each component included in Magnetic mRNA Isolation Kit and meet the designated specifications.	**Pass**
Functional Testing (mRNA Isolation) Oligo d(T)25 Magnetic Beads were equilibrated and incubated with freshly prepared eukaryotic cell lysate for direct mRNA isolation. The beads were washed and the mRNA eluted. The eluate was evaluated on an Agilent Bioanalyzer and the enriched poly(A)+ RNA contains ≤15% rRNA.	**Pass**

This product has been tested and shown to be in compliance with all specifications.

Michael Sproviero
Production Scientist
13 Jul 2020

Michael Tonello
Packaging Quality Control Inspector
13 Jul 2020

nqa ISO 9001 Registered Quality Management
nqa ISO 13485 Registered Medical Devices
nqa ISO 14001 Registered Environmental Management