

be INSPIRED drive DISCOVERY stay GENUINE

320 Newburyport Turnpike Rowley, MA 01969-2002 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Certificate of Analysis

Product Name:	Monarch® gDNA Purification Columns
Catalog Number:	T3017L
Packaging Lot Number:	10163996
Expiration Date:	12/2024
Storage Temperature:	25°C
Specification Version:	PS-T3017S/L v2.0

Monarch® gDNA Purification Columns Component List			
Component Description	Lot Number	Individual QC Result	
gDNA Purification 50 Columns	10170040	Pass	
Monarch® Collection Tubes II	10176509	Pass	
	Component Description gDNA Purification 50 Columns	Component DescriptionLot NumbergDNA Purification 50 Columns10170040	

Assay Name/Specification	Lot # 10163996
Functional Testing (DNA Recovery and Purity)	Pass
Twenty-four Monarch® gDNA Purification Columns are tested with 5 µg of HindIII	
digested Lambda resulting in ≥75% recovery in ≥80% of the samples. OD 260/280 and	
260/230 are ≥1.75 in ≥80% of the samples.	

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.

Ana Egana Production Scientist 05 Jul 2023

aum

June Gaumont Packaging Quality Control Inspector 05 Jul 2023

