240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Product Specification

Product Name: Ph.D.TM-12 Phage Display Peptide Library

Catalog #: E8111L

Concentration: $1 \times 10^{13} \text{ pfu/ml}$ Shelf Life: 24 months

Storage Temp: -20°C

Storage Conditions: 25 mM Tris-HCl, 75 mM NaCl, 50% Glycerol, (pH 7.5 @) 25°C)

Specification Version: PS-E8111L v2.0
Effective Date: 16 Apr 2019

Assay Name/Specification (minimum release criteria)

Absolute Phage Titer - Infection of a mid-log culture of *E. coli* ER2738 with Ph.D.TM-12 Phage Display Peptide Library followed by plating, yields $\geq 1 \times 10^{13}$ pfu/ml.

Functional Testing (Panning) - A 100-fold representation of the Ph.D.TM-12 Phage Display Peptide Library containing approximately 10^{11} pfu is diluted in 200 μ l TBS and panned against 300 ng β -endorphin monoclonal antibody. The bound phage is affinity captured using magnetic beads and eluted with 1 ml of 0.2M Glycine-HCl, pH 2.2. After three rounds of selection, \geq 75% of sequences contain a motif related to the known epitope for the antibody.

Phage Contamination (Environmental) - A 1:100 dilution of an overnight culture of *E. coli* ER2738 was made in 20 ml LB, to which 10³ pfu of Ph.D.TM-12 Phage Display Peptide Library was added. The flask was incubated at 37°C on a rotating shaker for 5 hours. A 1 ml volume of culture was removed and centrifuged. A volume of culture supernatant equivalent to the initial PFU input was added to a second, 20 ml culture like the first. The final culture supernatant was plated on three LB/IPTG/Xgal plates and then titered. Fewer than 20% clear or white plaques were observed in a minimum of 100 total plaques counted on each plate.

Sequence Verification (DNA) - The Ph.D.TM-12 Phage Display Peptide Library was sequenced using 5'-

CCCATGTACCGTAACACTGAGTTTC-3' as a primer to confirm the correct form of the cloned insert on the displayed peptide, X_{12} -GGG.

Luk Cotimon

Date 16 Apr 2019

Derek Robinson

Director of Quality Control





