

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Product Specification

Product Name:	SfoI
Catalog #:	R0606S/L
Concentration:	10,000 units/ml
Unit Definition:	One unit is defined as the amount of enzyme required to digest 1 μ g of Lambda DNA (HindIII digest) in rCutSmart Buffer in 1 hour at 37°C in a total reaction volume of 50 μ l.
Shelf Life:	24 months
Storage Temp:	-20°C
Storage Conditions:	10 mM Tris-HCl, 200 mM NaCl, 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 200 μg/ml rAlbumin (pH 7.4 @ 25°C)
Specification Version:	PS-R0606S/L v2.0
Effective Date:	05 Jun 2022

Assay Name/Specification (minimum release criteria)

Endonuclease Activity (Nicking) - A 50 μ l reaction in rCutSmartTM Buffer containing 1 μ g of supercoiled LITMUS28i DNA and a minimum of 30 units of SfoI incubated for 4 hours at 37°C results in <20% conversion to the nicked form as determined by agarose gel electrophoresis.

Exonuclease Activity (Radioactivity Release) - A 50 μ l reaction in rCutSmartTM Buffer containing 1 μ g of a mixture of single and double-stranded [³H] *E. coli* DNA and a minimum of 100 units of SfoI incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.

Ligation and Recutting (Terminal Integrity) - After a 10-fold over-digestion of Lambda-HindIII DNA with SfoI, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with SfoI.

Protein Purity Assay (SDS-PAGE) - SfoI is >95% pure as determined by SDS PAGE analysis using Coomassie Blue detection.

Non-Specific DNase Activity (16 Hour) - A 50 μ l reaction in rCutSmartTM Buffer containing 1 μ g of Lambda-HindIII DNA and a minimum of 30 units of SfoI incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit <u>www.neb.com/trademarks</u> for additional information.

Date 05 Jun 2022

Derek Robinson Director, Quality Control



PS-R0606S/L v2.0 Page 1 of 1