240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Product Specification

Product Name: Hpy188I

Catalog #: R0617S/L

Concentration: 10,000 units/ml

Unit Definition:

One unit is defined as the amount of enzyme required to digest 1 µg of pBR322 in 1 hour at 37°C in a total reaction volume of

50 μl.

Shelf Life: 24 months Storage Temp: $-20 \, ^{\circ}\text{C}$

Storage Conditions: 100 mM NaCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 200 µg/ml BSA

Specification Version: PS-R0617S/L v1.0

Effective Date: 05 Jun 2013

Assay Name/Specification (minimum release criteria)

Exonuclease Activity (Radioactivity Release) - A 50 μ l reaction in CutSmartTM Buffer containing 1 μ g of a mixture of single and double-stranded [3 H] *E. coli* DNA and a minimum of 30 units of Hpy188I incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.

Ligation and Recutting (Terminal Integrity) - After a 5-fold over-digestion of pBR322 DNA with Hpy188I, \sim 50% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, \sim 75% can be recut with Hpy188I.

Non-Specific DNase Activity (16 hour) - A 50 µl reaction in CutSmartTM Buffer containing 1 µg of pBR322 DNA and a minimum of 10 units of Hpy188I incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis. NOTE: although no nuclease degradation is detected under these conditions, extended incubations and/or high concentrations of this enzyme may result in star activity. See the product FAQ for recommended reaction conditions for this enzyme.

* The BSA in this product has been granted an EDQM "Certificate of Suitability" from the European Directorate for the Quality of Medicines (# R1-CEP-2003-204-Rev00) and has been granted a USDA Certificate for Export of Bovine Blood Plasma/Serum for Manufacture into Pharmaceutical Products.

Date

05 Jun 2013

Derek Robinson

Director of Quality Control

nga.
ISO 9001
Registered
Quality



