

New England Biolabs Certificate of Analysis

Product Name: *Taq PCR Kit*
 Catalog Number: *E5000S*
 Lot Number: *10048559*
 Expiration Date: *12/2020*
 Storage Temperature: *-20°C*
 Specification Version: *PS-E5000S v2.0*

Taq PCR Kit Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
N0550AVIAL	Quick-Load® Purple 1 kb Plus DNA Ladder	10033005	Pass
N0447SVIAL	Deoxynucleotide (dNTP) Solution Mix	10046884	Pass
M0273AVIAL	Taq DNA Polymerase with Standard Taq Buffer	10032556	Pass
B9021SVIAL	Magnesium Chloride (MgCl ₂) Solution	10038440	Pass
B9015SVIAL	Standard Taq (Mg-free) Reaction Buffer Pack	10033771	Pass
B9014SVIAL	Standard Taq Reaction Buffer Pack	10032877	Pass

Assay Name/Specification	Lot # 10048559
<p>PCR Amplification (5.0 kb Lambda DNA) A 50 µl reaction in Standard Taq Reaction Buffer in the presence of 200 µM dNTPs and 0.2 µM primers containing 5 ng Lambda DNA with 1.25 units of Taq DNA Polymerase for 25 cycles of PCR amplification results in the expected 5.0 kb product.</p>	Pass
<p>* Individual Product Component Note Standard Quality Control Tests are performed for each component included in Taq PCR Kit and meet the designated specifications.</p>	Pass

This product has been tested and shown to be in compliance with all specifications.



Tony Spear-Alfonso
Production Scientist
16 Jan 2019



Michael Tonello
Packaging Quality Control Inspector
23 Jul 2019