

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: NEBNext® Ultra™ II End Repair/dA-Tailing Module

Catalog Number: E7546S
Packaging Lot Number: 10071185
Expiration Date: 03/2021
Storage Temperature: -20°C

Specification Version: PS-E7546S/L v1.0

NEBNext® Ultra™ II End Repair/dA-Tailing Module Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
E7647AVIAL	NEBNext® Ultra™ II End Prep Reaction Buffer	10071187	Pass	
E7646AVIAL	NEBNext® Ultra™ II End Prep Enzyme Mix	10071186	Pass	

Assay Name/Specification	Lot # 10071185
Functional Testing (Library Construction, DNA Modules) Each set of reagents is functionally validated and compared to a previous lot through construction of DNA libraries. Libraries made from previous and current lots for the recommended input DNA amounts are sequenced together on the same Illumina® flow cell and compared across various metrics including library yield, fraction of reads aligning to the reference, GC bias, and insert size.	Pass
* Individual Product Component Note Standard Quality Control Tests are performed for each component included in NEBNext® Ultra™ II End Repair/dA-Tailing Module and meet the designated specifications.	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



E7546S / Lot: 10071185

Page 1 of 2

Christin Summ

Christine Sumner Production Scientist 27 Mar 2020 Michael Tonello

Packaging Quality Control Inspector

04 Aug 2020

E7546S / Lot: 10071185

Page 2 of 2