

be INSPIRED *drive* DISCOVERY *stay* GENUINE

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name:	NEBNext® Ultra™ II Directional RNA Second Strand Synthesis Module
Catalog Number:	E7550S
Packaging Lot Number:	10097125
Expiration Date:	10/2021
Storage Temperature:	-20°C
Specification Version:	PS-E7550S/L v1.0

NEBNext® Ultra [™] II Directional RNA Second Strand Synthesis Module Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
	NEBNext® Second Strand Synthesis Reaction Buffer (dUTP Mix)	10071773	Pass	
E7425AVIAL	NEBNext® Second Strand Synthesis Enzyme Mix	10071772	Pass	

Assay Name/Specification	Lot # 10097125
Functional Testing (Library Construction, RNA) Each set of reagents is functionally validated and compared to the previous lot through construction of libraries made from commercially available RNA, using the kit's minimum and maximum input requirements. Libraries made from the previous and current lots for both input RNA amounts are sequenced together on the same Illumina flow cell and compared across various metrics including library yield, individual transcript abundance correlations (low vs. high input, old lot vs. new lot), 5'-3' transcript coverage, and fraction of reads mapping to a reference.	Pass
* Individual Product Component Note Standard Quality Control Tests are performed for each component included in NEBNext® Ultra™ II Directional RNA Second Strand Synthesis Module and meet the designated specifications.	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.





be INSPIRED drive DISCOVERY stay GENUINE

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

notin-

Christine Sumner Production Scientist 13 Jan 2021

Michae on. V

Michael Tonello Packaging Quality Control Inspector 13 Jan 2021

