

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Certificate of Analysis

Product Name: DNA Polymerase I (E. coli)

Catalog Number: M0209L Concentration: 10,000 U/ml

Unit Definition: One unit is defined as the amount of enzyme that will incorporate 10

nmol of dNTP into acid insoluble material in 30 minutes at 37°C.

Lot Number: 10051040
Expiration Date: 06/2021
Storage Temperature: -20°C

Storage Conditions: 25 mM Tris-HCl , 1 mM DTT , 0.1 mM EDTA , 50 % Glycerol, (pH 7.4 @

25°C)

Specification Version: PS-M0209S/L v1.0

DNA Polymerase I (E. coli) Component List				
<b>NEB Part Number</b>	Component Description	Lot Number	Individual QC Result	
M0209LVIAL	DNA Polymerase I (E. coli)	10044670	Pass	
B7002SVIAL	NEBuffer™ 2	10043903	Pass	

Assay Name/Specification	Lot # 10051040
Endonuclease Activity (Nicking) A 50 μl reaction in NEBuffer 2 containing 1 μg of supercoiled PhiX174 DNA and a minimum of 100 units of DNA Polymerase I (E. coli) incubated for 4 hours at 37°C results in <10% conversion to the nicked form as determined by agarose gel electrophoresis.	Pass
Phosphatase Activity (pNPP) A 200 µl reaction in 1M Diethanolamine, pH 9.8, 0.5 mM MgCl2 containing 2.5 mM p-Nitrophenyl Phosphate (pNPP) and a minimum of 100 units DNA Polymerase I (E. coli) incubated for 4 hours at 37°C yields <0.0001 unit of alkaline phosphatase activity as determined by spectrophotometric analysis.	Pass
Protein Purity Assay (SDS-PAGE)  DNA Polymerase I (E. coli) is ≥ 99% pure as determined by SDS-PAGE analysis using Coomassie Blue detection.	Pass
qPCR DNA Contamination (E. coli Genomic) A minimum of 10 units of DNA Polymerase I (E. coli) is screened for the presence of E. coli genomic DNA using SYBR® Green qPCR with primers specific for the E. coli 16S	Pass



M0209L / Lot: 10051040

Page 1 of 2

gel electrophoresis using fluorescent detection.

This product has been tested and shown to be in compliance with all specifications.

Production Scientist

17 Jun 2019

Michael Tonello

Packaging Quality Control Inspector

01 Aug 2019

