

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: GpC Methyltransferase (M.CviPI)

Catalog Number: M0227L
Concentration: 4,000 U/ml

Unit Definition: One unit is defined as the amount of enzyme required to protect 1 µg

Lambda DNA in 1 hour at 37°C in a total reaction volume of 20 μl

against cleavage by HaeIII restriction endonuclease.

Packaging Lot Number: 10164704
Expiration Date: 12/2023
Storage Temperature: -20°C

Storage Conditions: 15 mM Tris-HCl, 200 mM NaCl, 1 mM DTT, 0.1 mM EDTA, 50 % Glycerol,

200 $\mu g/ml$ BSA, (pH 7.4 @ 25°C)

Specification Version: PS-M0227S/L v2.0

GpC Methyltransferase (M.CviPI) Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
M0227LVIAL	GpC Methyltransferase (M.CviPI)	10132530	Pass
B9003SVIAL	S-adenosylmethionine (SAM)	10153874	Pass
B0227SVIAL	GC Reaction Buffer	10148519	Pass

Assay Name/Specification	Lot # 10164704
Non-Specific DNase Activity (16 Hour) A 50 μl reaction in GC Reaction Buffer containing 1 μg of Lambda DNA and a minimum of 40 units of GpC Methyltransferase (M.CviPI) incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
Functional Testing (Methyltransferase) A 20 μl reaction in GC Reaction Buffer supplemented with 160 μM SAM containing 1 μg of Lambda DNA and 1 unit of GpC Methyltransferase (M.CviPI) incubated for 1 hour at 37°C followed by heat inactivation results in ≥ 95% protection from digestion with 10 units of HaeIII in NEBuffer 2 incubated at 37°C for 1 hour as determined by agarose gel electrophoresis.	Pass
Exonuclease Activity (Radioactivity Release) A 50 μl reaction in GC Reaction Buffer containing 1 μg of a mixture of single and double-stranded [³H] E. coli DNA and a minimum of 40 units of GpC Methyltransferase (M.CviPI) incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass



M0227L / Lot: 10164704

Page 1 of 2

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.

Tim Meixsell Production Scientist

05 Jan 2022

Josh Hersey

Packaging Quality Control Inspector

11 Oct 2022

