

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: LongAmp® Taq DNA Polymerase

Catalog Number: M0323S
Concentration: 2,500 U/ml

Unit Definition: One unit is defined as the amount of enzyme that will incorporate 10

nmol of dNTP into acid insoluble material in 30 minutes at 75°C.

Packaging Lot Number: 10069145
Expiration Date: 08/2021
Storage Temperature: -20°C

Storage Conditions: 10 mM Tris-HCl , 100 mM KCl , 1 mM DTT , 0.1 mM EDTA , 0.5 % Tween®

20 , 0.5 % IGEPAL® CA-630 , 50 % Glycerol, (pH 7.4 @ 25°C)

Specification Version: PS-M0323S/L v2.0

LongAmp® Taq DNA Polymerase Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
M0323SVIAL	LongAmp® Taq DNA Polymerase	10049392	Pass	
B0323SVIAL	LongAmp® Taq Reaction Buffer	10042779	Pass	

Assay Name/Specification	Lot # 10069145
RNase Activity (Extended Digestion) A 10 µl reaction in NEBuffer 4 containing 40 ng of a 300 base single-stranded RNA and a minimum of 1 µl of LongAmp® Taq DNA Polymerase is incubated at 37°C. After incubation for 16 hours, >90% of the substrate RNA remains intact as determined by gel electrophoresis using fluorescent detection.	Pass
PCR Amplification (30 kb Lambda DNA) A 25 μl reaction in LongAmp® Taq Reaction Buffer in the presence of 300 μM dNTPs and 0.4 μM primers containing 1 ng Lambda DNA with 2.5 units of LongAmp® Taq DNA Polymerase for 28 cycles of PCR amplification results in the expected 30 kb product.	Pass
PCR Amplification (30 kb Human Genomic DNA) A 25 μl reaction in LongAmp® Taq Reaction Buffer in the presence of 300 μM dNTPs and 0.4 μM primers containing 500 ng Human Genomic DNA with 2.5 units of LongAmp® Taq DNA Polymerase for 28 cycles of PCR amplification results in the expected 30 kb product.	Pass
qPCR DNA Contamination (E. coli Genomic) A minimum of 2.5 units of LongAmp® Taq DNA Polymerase is screened for the presence	Pass



M0323S / Lot: 10069145 Page 1 of 2 This product has been tested and shown to be in compliance with all specifications.

detectable nuclease degradation as determined by agarose gel electrophoresis.

Christie Vazquez
Production Scientist

16 Apr 2020

ay Minichiello

Packaging Quality Control Inspector

16 Apr 2020

