

240 County Road Ipswich, MA 01938-2723

Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Certificate of Analysis

Product Name: Q5® High-Fidelity 2X Master Mix

Catalog Number: M0492S

NEW ENGLAND

Concentration: 2 X Concentrate

Packaging Lot Number: 10060946
Expiration Date: 10/2021
Storage Temperature: -20°C

Specification Version: PS-M0492S/L v1.0

Composition (1X): Proprietary

| Q5® High-Fidelity 2X Master Mix Component List |                                 |            |                      |  |
|------------------------------------------------|---------------------------------|------------|----------------------|--|
| <b>NEB Part Number</b>                         | Component Description           | Lot Number | Individual QC Result |  |
| M0492SVIAL                                     | Q5® High-Fidelity 2X Master Mix | 10051982   | Pass                 |  |

| Assay Name/Specification                                                                                                                                                                                                                                                                                                                          | Lot # 10060946 |
|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|----------------|
| Endonuclease Activity (Nicking, Polymerase, dNTP) A 50 μl reaction in NEBuffer 2 in the presence of 400 μM dNTPs containing 1 μg of supercoiled pUC19 DNA and a minimum of 10 units of Q5® High-Fidelity DNA Polymerase incubated for 4 hours at 37°C results in <10% conversion to the nicked form as determined by agarose gel electrophoresis. | Pass           |
| Non-Specific DNase Activity (16 hour, Buffer) A 50 µl reaction in 1X Q5® High-Fidelity Master Mix containing 1 µg of T3 DNA in addition to a reaction containing Lambda-HindIII DNA incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.                 | Pass           |
| PCR Amplification (20 kb Lambda DNA, Master Mix) A 50 µl reaction in 1X Q5® High-Fidelity Master Mix and 1.0 µM primers containing 10 ng Lambda DNA for 22 cycles of PCR amplification results in the expected 20 kb product.                                                                                                                     | Pass           |
| PCR Amplification (7 kb Human Genomic DNA, Master Mix) A 50 µl reaction in 1X Q5® High-Fidelity Master Mix and 0.5 µM primers containing 20 ng Human Genomic DNA for 30 cycles of PCR amplification results in the expected 7 kb product.                                                                                                         | Pass           |
| Phosphatase Activity (pNPP)                                                                                                                                                                                                                                                                                                                       | Pass           |



M0492S / Lot: 10060946

Page 1 of 2

| Assay Name/Specification                                                                                                                                                                                                                                                                                                                                                                                            | Lot # 10060946 |  |
|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|----------------|--|
| A 200 µl reaction in 1M Diethanolamine, pH 9.8, 0.5 mM MgCl2 containing 2.5 mM p-Nitrophenyl Phosphate (pNPP) and a minimum of 100 units of Q5® High-Fidelity DNA Polymerase incubated for 4 hours at 37°C yields <0.0001 unit of alkaline phosphatase activity as determined by spectrophotometric analysis.                                                                                                       |                |  |
| Protein Purity Assay (SDS-PAGE) Q5® High-Fidelity DNA Polymerase is ≥ 95% pure as determined by SDS-PAGE analysis using Coomassie Blue detection.                                                                                                                                                                                                                                                                   | Pass           |  |
| <b>qPCR DNA Contamination (E. coli Genomic)</b> A minimum of 2 units of Q5® High-Fidelity DNA Polymerase is screened for the presence of E. coli genomic DNA using SYBR® Green qPCR with primers specific for the E. coli 16S rRNA locus. Results are quantified using a standard curve generated from purified E. coli genomic DNA. The measured level of E. coli genomic DNA contamination is ≤ 1 E. coli genome. | Pass           |  |
| RNase Activity (Extended Digestion) A 10 µl reaction in NEBuffer 4 containing 40 ng of a 300 base single-stranded RNA and a minimum of 1 µl of Q5® High-Fidelity 2X Master Mix is incubated at 37°C. After incubation for 4 hours, >90% of the substrate RNA remains intact as determined by gel electrophoresis using fluorescent detection.                                                                       | Pass           |  |

This product has been tested and shown to be in compliance with all specifications.

Doreen Duquette **Production Scientist** 

04 Oct 2019

Josh Hersey Packaging Quality Control Inspector 05 Dec 2019



M0492S / Lot: 10060946

Page 2 of 2