

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: pBR322 DNA-Mspl Digest

Catalog Number: N3032S
Concentration: 1,000 µg/ml

Unit Definition: N/A

Lot Number: 10051120
Expiration Date: 08/2021
Storage Temperature: -20°C

Storage Conditions: 10 mM Tris-HCl (pH 8.0), 1 mM EDTA

Specification Version: PS-N3032S/L v1.0

pBR322 DNA-Mspl Digest Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
N3032SVIAL	pBR322 DNA-Mspl Digest	10051119	Pass	
B7025SVIAL	Gel Loading Dye, Purple (6X), no SDS	10038961	Pass	

Assay Name/Specification	Lot # 10051120
A260/A280 Assay The ratio of UV absorption of pBR322 DNA-Mspl Digest at 260 and 280 nm is between	Pass
1.8 and 2.0.	
DNA Concentration (A260)	Pass
The concentration of pBR322 DNA-Mspl Digest is between 1000 and 1050 μg/ml as determined by UV absorption at 260 nm.	
acternation by a value of pitch at 200 min.	
Electrophoretic Pattern (Marker)	Pass
The banding pattern of pBR322 DNA-Mspl Digest on a 3% agarose gel shows discrete, clearly identifiable bands at each band of the marker, when stained with Ethidium	
Bromide at a concentration of 0.5 µg/ml.	
Non-Specific DNase Activity (DNA, 16 hour)	Pass
A 50 µl reaction in 1X NEBuffer 2 containing 5 µg of pBR322 DNA-Mspl Digest	
incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease	
degradation as determined by agarose gel electrophoresis.	

This product has been tested and shown to be in compliance with all specifications.



N3032S / Lot: 10051120

Page 1 of 2



Production Scientist 19 Apr 2019 Michael Tonello

Packaging Quality Control Inspector

27 Aug 2019