NEW ENGLAND BioLabs® Inc.

240 County Road
Ipswich, MA 01938-2723

Tel 978-927-5054
Fax 978-921-1350

www.neb.com
info@neb.com

New England Biolabs
Certificate of Analysis

Product Name:	*SwaI*
Catalog #:	*R0604S/L*
Concentration:	*10,000 units/ml*
Unit Definition:	*One unit is defined as the amount of enzyme required to digest 1 µg of pXba DNA in 1 hour at 25°C in a total reaction volume of 50 µl.*
Lot #:	*0031703*
Assay Date:	*03/2017*
Expiration Date:	*3/2019*
Storage Temp:	*-20°C*
Storage Conditions:	*400 mM NaCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 200 µg/ml BSA*
Specification Version:	*PS-R0604S/L v2.0*
Effective Date:	*30 Mar 2016*

Assay Name/Specification (minimum release criteria)	**Lot #0031703**
Exonuclease Activity (Radioactivity Release) - A 50 µl reaction in NEBuffer 3.1 containing 1 µg of a mixture of single and double-stranded [^{3}H] *E. coli* DNA and a minimum of 100 units of SwaI incubated for 4 hours at 25°C releases <0.1% of the total radioactivity.	**Pass**
Ligation and Recutting (Terminal Integrity) - After a 20-fold over-digestion of pXba-NdeI DNA with SwaI, ~75% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, ~75% can be recut with SwaI.	**Pass**
Non-Specific DNase Activity (16 Hour) - A 50 µl reaction in NEBuffer 3.1 containing 1 µg of pXba DNA and a minimum of 100 units of SwaI incubated for 16 hours at 25°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	**Pass**

* *The BSA in this product has been granted an EDQM "Certificate of Suitability" from the European Directorate for the Quality of Medicines (# R1-CEP-2003-204-Rev00) and has been granted a USDA Certificate for Export of Bovine Blood Plasma/Serum for Manufacture into Pharmaceutical Products.*

Authorized by
Derek Robinson
30 Mar 2016

Inspected by
Penghua Zhang
05 Apr 2017

ISO 9001 Registered Quality Management
ISO 14001 Registered Environmental Management
ISO 13485 Registered Medical Devices