


## New England Biolabs Certificate of Analysis

Product Name: *Monarch RNA Cleanup Columns (10 µg)*  
 Catalog Number: *T2037L*  
 Lot Number: *0011806*  
 Expiration Date: *12/2019*  
 Storage Temperature: *25°C*  
 Specification Version: *PS-T2037L v1.0*

Assay Name/Specification	Lot # 0011806
<p><b>Functional Testing (RNA Recovery and Purity)</b>            Twenty-four Monarch® RNA Cleanup Columns (T2037, 10 µg) are tested with 1 µg of rRNA resulting in ≥75% recovery in ≥80% of the samples. OD 260/280 and 260/230 are ≥1.75 in ≥80% of the samples.</p>	<p><b>Pass</b></p>

This product has been tested and shown to be in compliance with all specifications.



Ana Egana  
Production Scientist  
31 Aug 2018



Heidi Church  
Packaging Quality Control Inspector  
31 Aug 2018