

## New England Biolabs Certificate of Analysis

**Product Name:** *Monarch® gDNA Wash Buffer*  
**Catalog Number:** *T3015L*  
**Packaging Lot Number:** *10231527*  
**Expiration Date:** *03/2026*  
**Storage Temperature:** *25°C*  
**Specification Version:** *PS-T3015S/L v2.0*

Monarch® gDNA Wash Buffer Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
T3015-2	Monarch® gDNA Wash Buffer	10231309	Pass

Assay Name/Specification	Lot # 10231527
<b>Functional Testing (DNA Recovery and Purity)</b> Twenty-four Monarch® gDNA Purification Columns are tested with 5 µg of HindIII digested Lambda resulting in ≥75% recovery in ≥80% of the samples. OD 260/280 and 260/230 are ≥1.75 in ≥80% of the samples.	<b>Pass</b>

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit [www.neb.com/trademarks](http://www.neb.com/trademarks) for additional information.



Nikki Samuelson  
Production Scientist  
15 Mar 2024



Julie Gaumont  
Packaging Quality Control Inspector  
20 Mar 2024